


# HELMET SIZING AND SELECTION INFORMATION



Since proper fit is important for the effectiveness of the helmet as well as comfort of the person wearing headgear, measurements should be determined by professional personnel.

Measure the head circumference (all the way around) (A) at the eyebrow level. The Occipital measurement (B) is taken from the eyebrow level to the back of the head at the point that the helmet is to end. For (C) measure from the top of one ear over the head to the top of the other ear. If for some reason the measurements do not fall within the above guidelines, or if there are other considerations to be taken into account, please refer to the next page for information on Custom Helmets

(A) CIRCUMFERENCE RANGE

	Soft Shell	Hard Shell
Infant	16 ¼ - 17	Not available in infant
XX-Small	17 ½ - 18 ¼	18½ - 19 ¼
X-Small	19 - 19 ¾	19 ¼ - 20
Small	20 ¼ - 21 ½	20 ½ - 21 ¼
Medium	21 ½ - 22 ¼	21 ½ - 22 ¼
Large	23 - 23 ¾	22 ½ - 23 ¼
X-Large	24 - 24 ¾	24 ½ - 25 ¼
XX-Large	25 ½ - 26 ¼	26" - 26 ¼

(B) OCCIPITAL

	Soft Shell	Hard Shell
Infant	11 ½	Not available in infant
XX-Small	12 ¾	12 ¾
X-Small	13 ¾	13 ¾
Small	14 ½	15
Medium	15 ¼	16 ½
Large	16 ½	16 ¾
X-Large	17 ½	17 ½
XX-Large	18 ½	18 ¾

(C) EAR TO EAR

	Soft Shell	Hard Shell
Infant	8 ¼	Not available in infant
XX-Small	9 ½	9 ¼
X-Small	9 ½"	10
Small	10 ½	11 ¾
Medium	11	12 ½
Large	12	12 ¾
X-Large	13	13 ¼
XX-Large	13 ¾	15

Measurements are in inches

## HEADGEAR SELECTION

Some suggested considerations to use in helping to determine that type of helmet would be suitable for your client are:

- Degree of physical activity.
- Potential for striking surfaces (floor, walls, furniture, etc.).
- Nature of self-abusiveness.
- Severity and frequency of seizures or spasticity.
- Length of time worn.

**CUSTOM AND/OR MODIFIED ITEMS ARE NOT RETURNABLE.**