Posey Transfer and Walking Belts

Posey Transfer and Walking Belts offer caregivers added support with multiple easy to grasp hand loops. The belt is made of machine washable black nylon webbing with green nylon hand loops. Belts are available with nickel-plated or guick-release buckle closures.





6537QDX

Ref #	Fits Waists	No. Of Handles	Buckle	Rear Support Band
REF 6537DX	30" to 66" (76 cm to 168 cm)	7	Metal	6" (15 cm) wide
REF 6537QDX	30" to 66" (76 cm to 168 cm)	7		6" (15 cm) wide
REF 6537	28" to 52" (71 cm to 132 cm)	6		4" (10 cm) wide
REF 6537Q	28" to 52" (71 cm to 132 cm)	6		4" (10 cm) wide

Application Instructions for Metal and Quick-Release Buckle:

Follow standard instructions on previous pages for proper application of metal and guick-release buckles.

Washing Instructions for Posey Cotton and/or Polyester Transfer and Gait Belts:

For maintaining bright colors in **non-contaminated Belts:**

160°F

For all **contaminated** Gait/Transfer Belts:

WASH HOT BLEACH AS DIRECTED DRY ON

NOTE: Fasten all buckles to minimize damage during wash and dry cycles. Close hook and loop fasteners before washing to prevent lint build-up. Use stiff metal brush to clean "hook" material after washing. Test hook and loop adhesion before use, and discard belt if hook and loop does not hold securely. During the washing, process, metal buckles could bend and guick-release buckles could crack or break. DO NOT put buckles through extractors. ALWAYS ensure that all buckles are in good working order and are not cracked or broken before each use. Customers could expect up to 5% shrinkage in cotton materials over the life of the product.

Disposal

AWARNING Properly dispose of the product per facility's policy for BIOHAZARDOUS materials.

Storage and Handling

This device is designed for use in normal indoor environments. This device may be stored in ambient warehouse temperatures at normal humidity levels. Avoid excess moisture or high humidity that may damage product materials.

Other Posey Gait Belts

Improve Your Walk-in Business with Posey Retail Packaged Gait Belts.

Most Posey Gait Belts are available in display retail packaging at a slightly higher price. Call Customer Service at 1.800.447.6739 for ordering information.

- Attractive see-through packaging.
- Easily hangs from peg hooks or can be placed on a shelf.
- Increased product display visibility.
- Clamshell package easily opens and reseals for product examination.
- Available in a wide assortment of Posey Gait Belts.

Posey Premium EZ Clean Gait Belts

Laminated webbing wipes clean to help address facility cross-contamination concerns. Spring loaded, self-locking buckle is easy to apply and molded from rugged polyethylene. 2" (5 cm) wide webbing. Available in standard and bariatric lengths and in five colors.

Webbing	Fits Waists Up To			
Color	58"L (147 cm)	70"L (178 cm)		
Pink	6546P	6546PL		
Black	6546	6546L		
Blue	6546B	6546BL		
Orange	6546O	6546OL		
Yellow	6546Y	6546YL		

Posey Add-On Gait Belt Handles

Upgrade standard Gait Belts with a vertical hand grip for better support when assisting patients. Fits all 2" wide Gait Belts and the ergonomically contoured soft touch, polymer handles provide a firmer handhold. Weight capacity of 300 lbs. (136 kgs.) per handle.

Ref. # Description Qty. **REF 6558** Posey Handles 1 pr.







Posey Gait and Transfer Belts

Posev Gait Belts provide caregivers with a secure hold point while assisting patients during supervised walking and transfers. When used correctly, they may also help to prevent back injuries to the caregiver. When not in use, the gait belt can be worn by the caregiver to ensure immediate availability when needed.

_	
Avai	ilable Styles
Whit	e Cotton
Cora	l Reef
Conf	
Lave	nder
Fiest	a
	Blue
Sout	hwest Embroidered
Bouc	juet - Heavy Duty Webbi
Bouc	
E-Z C	Clean Black 60" (152 cm)
	Clean Heavy Duty Black 6
	Clean Heavy Duty Blue 60
E-Z C	Clean Heavy Duty Pink 60
	Clean Heavy Duty Orange
	Ilean Heavy Duty Yellow
Whit	e
Navy	
Bouc	
	e Navy 62" (157 cm)
	e White 62" (157 cm)
	and Stripes - Ribbon
	onal Spring - Tulips
	onal Summer - Beach
	onal Autumn - Leaves
Seaso	onal Winter - Snowmen
N	OTE: Standard length is 54" (13

The Proper Use of Posey Gait & Transfer

Gait and Transfer Belts

Indications: Patients/Residents requiring ambulation and/or transfer assistance.



	BUCKLE TYPE/CLOSURE			URE	REORDER NUMBER		
	NICKEL- Plated	QUICK- Release	AIRLINE	SPRING	HOOK AND LOOP	STANDARD (See Note)	EXTRA LONG 70" (178 cm)
ng 0" (152 cm) " (152 cm) " (152 cm) 60" (152 cm) 60" (152 cm)	••••••	• • •	•	$\bullet \bullet \bullet \bullet$	• •	 □ 6524 □ 6525 □ 65251 □ 65252 □ 6527 □ 6528 □ 6529 □ 6530 □ 6545 □ 6545 □ 6546 □ 65466 □ 65466 □ 65466 □ 65466 □ 65466 □ 65462 □ 6524Q □ 6528Q □ 6528Q □ 6528V □ 6528V □ 6528V □ 65558 □ 65555 □ 65555 □ 65555 □ 65555 	 6524L 6525L 65251L 65251L 6527L 6528L 6529L 6530L 6530L 6536L 6546BL 6546BL 6546PL 65546PL 6555LSP 6555LF 6555LW

NOTE: Standard length is 54" (137 cm) unless otherwise noted. Colors and patterns on belts subject to change without notice.

Inservice Video Available!

The Proper Use of Posey Gait and Transfer Belts (Cat. 8508) is a 15 minute video that demonstrates the proper application of Posey Transfer and Walking Belts. It reviews nine basic rules for successful transfers and three common transfer procedures.

NOTE: THESE PRODUCTS ARE NOT INTENDED TO BE USED AS A RESTRAINT!

The Posey Company recommends the user of this product check and follow its facility's written procedures for approved assisted walking, lifting, and/or transfer procedures.

Posey Company • 5635 Peck Road, Arcadia, CA 91006-0020 USA • www.posey.com Phone: 1.800.447.6739 or 1.626.443.3143 • Fax: 1.800.767.3933 or 1.626.443.5014

Contraindications include, but are not limited to:

Patients with ostomy, colostomy, G-tubes, Hernias, severe Cardio Obstructive Pulmonary Disease (COPD), as well as those with post-surgery incisions, or fitted with monitoring equipment, tubes or lines that might be compromised by the pressure or the rubbing against a Gait Belt. NEVER apply a Gait Belt if there is any compromise of safety or indication of inappropriateness to the patient or caregiver.

Monitoring

AWARNING Be sure to follow your facility's policies and guidelines for frequency of patient monitoring. Check for skin integrity, proper circulation and range of motion and rehabilitation. ALWAYS test to make sure the belt is applied securely and will not loosen or give when transferring patient.

Inspection

AWARNING Inspect before each use: check for broken stitches or parts; torn, cut or frayed material: or buckles that are cracked or broken and do not hold securely. DO NOT use soiled or damaged products.

Application Instructions for Metal Buckle:

- **1.** With the Posey label facing outside, wrap the belt around the patient's waist, pass the metal tip of the belt through the buckle and over the teeth of the buckle (Fig. 1).
- 2. Bring the tip of the belt across the front of the buckle and slip it through the other side (Fig. 2). Adjust it so it is snug, but not uncomfortable for the patient. You should be able to slip your open hand (flat) between the belt and the patient.
- 3. The strap should lay flat across the buckle. Tuck excess under the belt (Fig. 3). ALWAYS verify proper closure before use.
- 4. ALWAYS grasp the transfer belt from underneath (Fig. 4). Assist patient using your facility's approved procedures.










Application Instructions for Quick-Release Buckle:

- 1. Release the quick-release buckle by pushing inward on the quickrelease buttons.
- 2. Fit the belt to the patient's waist size
- 3. Once adjusted, wrap the belt around the patient and snap the buckle shut.
- 4. Make sure you can slide your open hand (flat) between the belt and the patient.
- 5. Turn the buckle around to the patient's back (tuck excess strap under belt).

AWARNING The guick-release buckle should be positioned so that the patient can not release the buckle during transfer. Release of the buckle during transfer may result in injury. ALWAYS verify the buckle is not broken or cracked and holds securely before transferring the patient.

Application Instructions for the Posey E-Zee closure:

- 1. With the Posey label facing away from the patient, wrap the belt around the patient's waist and pass the end through the first metal ring on the belt.
- **2.** For waists up to 42" (107 cm), pass the tip through the second metal ring also. For waists 43" to 54" (109 cm to 137 cm), use only the first metal ring.
- 3. After passing the belt through the appropriate metal ring(s). fold it over itself and back in the opposite direction.
- 4. Secure the hook and loop closure by pressing both sides together firmly.
- 5. ALWAYS verify proper closure before use. Adjust the belt so it is snug, but not uncomfortable for the patient. You should be able to slide your open hand (flat) between the belt and the patient.



AWARNING NEVER use a Posey product as a seat belt in a moving vehicle. Posey products are not

designed to withstand the force of a crash.

Small

Large

REF 6534XL X-Large

Medium

Application Instructions:

REF 6534S

BEF 6534M

REF 6534L

- **1.** Center the belt around the patient's waist with the quickrelease buckles in the center front and the rear adjustment strap at the top of the belt.
- **2.** Adjust the belt to fit so it is snug, but not uncomfortable for the patient, and close it with the hook and loop fastener. You should be able to slide your open hand (flat) between the belt and the patient.
- 3. Secure the guick-release buckles and tighten the straps so that they lie flat.
- 4. Tighten the cotton strap in the rear of against the belt.



Posey Ergonomic Walking Belt


This white cotton belt is 5" (13 cm) wide and has three hand loops, one at each side and one horizontal loop in the back. The side handles are reinforced with plastic for a more secure grip. Attaches quickly with plastic quick-release buckles. Adjustable strap in back for improved fit. Machine washable.



FRONT

SIZING CHART

S: 26"-34" (66 cm to 86 cm)	L: 36"-45" (91 cm to 114 cm)
M: 33"-38" (84 cm to 97 cm)	XL: 44"-52" (112 cm to 132 cm)



the belt until it fits snugly in the waist area. Make sure the strap and buckle lie flat

BACK