




UNPACKING THE DEVICE

- (1) Brace
- (2) Eyelet Hooks
- (2) 5/16" Screws
- (2) 1/4" Screws
- (2) 3/8" Screws
- (2) Washers



PARTS OF THE BRACE



1 OPEN ANKLE CUFF



Too loosen the Ankle Cuff Cord, turn Ankle Cuff Dial counterclockwise.



Grab Pull Tab and pull until the Magnetic Fastener becomes disconnected from the base.

2 POSITION THE DEVICE



Place the SaeboStep around the ankle with the Front Dial centered to your shoe and the blue Saebo logo positioned at the top.

3 SECURE ANKLE CUFF



Grab Pull Tab that is attached to the end of the Ankle Cuff Cord and connect to magnetic base. You will hear a click when secured.



Turn Ankle Cuff Dial clockwise until the Cuff is fit snugly.

SECURE ANKLE CUFF CONTINUED



Note 1: Make sure the Ankle Cuff Cord is not twisted.



Note 2: The Cuff on the dial side should overlap the Cuff on the magnetic side.

4 ATTACH EYELET HOOKS



Insert proper Screw and Washer into shoe eyelet.



Stabilize the head of the screw with your finger while securely twisting the Eyelet Hook with other hand. Position the open end of the Eyelet Hook toward the front of the shoe.

5 LOOSEN FRONT CORD



Turn Front Dial counterclockwise.



Pull left and right Front Cords, ideally at the same time, downward toward the shoe.

6 ATTACH CORD TO EYELET HOOKS



Attach Cord onto the Eyelet Hook openings.

7 TIGHTEN FRONT CORD



Wind the dial by turning clockwise until the Front Cord is tight and you have the lift and angle desired.

USING ONE HAND



- 1 Rotate Ankle Cuff so Magnetic Fastener and Ankle Cuff Dial are near the front for easier access.
- 2 Attach Magnetic Fastener to Base.
- 3 Tighten Ankle Cuff Cord by winding the Ankle Cuff Dial clockwise.

- 4 Rotate brace back to the proper position with the Saebo logo facing the front.
- 5 Finish tightening the Ankle Cuff Cord by winding the Ankle Cuff Dial clockwise until you feel a snug fit.



WARNING
This product can expose you to: Chemicals of Potential Concern.
For more information go to: www.P65Warnings.ca.gov